

Summary of SIR Subcommittee Meeting

July 23, 2019

Present: Judy Morgan, Gail Warren, Carl Kircher, Harold Longbaugh, Lynn Bradley

Participants discussed the final four draft Implementation Guidance documents from New Orleans.

For the Chlorine Residual IG, the references should be changed to eliminate the V1M1/PT reference and to replace the V1M4 language with V1M5 language.

The New Scope Old Method IG is ready to post, as is.

The New Calibration Curve Overwritten IG needs revision to explain that before reprocessing all data, the old and new calibration curves should be compared to see if there is any difference to warrant re-running the samples. Judy agreed to make these edits.

The PT IG is ready to post, as is.

SIR 357 was discussed, about criteria for what is considered to be a “mechanical device”. This question was determined not to be a valid SIR, but the term “mechanical device” was forwarded to CSDEC Chair and Program Administrator for possible addition to the TNI Glossary. In the end, since the glossary is a “living document” and as such, not enforceable by regulatory agencies, the term will not be added.

The next SIR Subcommittee meeting will be Tuesday, August 27, 2019, at 12:30 pm Eastern.