

Summary of SIR Subcommittee Meeting

August 27, 2019

Present: Judy Morgan, Gail Warren, Carl Kircher, Harold Longbaugh, Lynn Bradley with Bill Hall sending approval of SIR350 response by email.

The response to SIR 350 was returned from Quality Systems, to say that all three conditions must be met in V1M2 5.2.6.2.c in order for the exception to apply. There was no adverse comment and this SIR will be posted to the NELAP AC's voting site.

Participants discussed the remaining two draft Implementation Guidance documents that were revised after the July meeting.

For the Chlorine Residual IG, the references were changed and this is ready to post to the website.

The New Calibration Curve Overwritten IG needs further revision to note that V1M4 does require use of the "most current" calibration curve, and to add an option for lab-specified procedures. Judy agreed to make these edits, and participants agreed that her changes would be acceptable without further review.

The next SIR Subcommittee meeting will be Tuesday, September 24, 2019, at 12:30 pm Eastern.